

Scatter Correction for Cone-beam CT in Radiation Therapy

A scatter correction scheme for X-ray CBCT based on CT images in the radiation treatment planning, which has the potential of outperforming all competing scatter correction techniques

Georgia Tech inventors are developing a new scatter correction scheme for X-ray CBCT based on the routinely used diagnostic multi-detector CT (MDCT) images in the radiation treatment planning, which has the potential of outperforming all competing scatter correction techniques. This technology demonstrates that the scatter signal can be obtained using the planning MDCT data due to its 'free' prior information.

To realize the scatter removal scheme, the primary signals of the CBCT projects are simulated via forward projection on MDCT data. The subtraction between the simulated projections from the true CBCT projection provides the scatters signals plus the primary signal difference between the MDCT and the CBCT scans, which is mainly due to the patient geometry deformation. After an accurate patient setup, the deformation only leads to high frequency primary signal perturbations. Based on the dominant-low frequency behavior of scatter signals, accurate scatter estimations are finally achieved via smoothing techniques.

Summary Bullets

- Does not modify either the hardware or the imaging protocols of the clinical system
- Effectively reduces the scatter signals in the CBCT scan, and improves image quality
- Does not require an increase in the dose to the patients

Solution Advantages

- Does not modify either the hardware or the imaging protocols of the clinical system
- Effectively reduces the scatter signals in the CBCT scan, and improves image quality
- Does not require an increase in the dose to the patients
- Does not require an increase in scan time
- While merits of the CBCT are retained, the new method utilizes planned MDCT as information to enhance the quality of the CBCT images
- High computation efficiently

Potential Commercial Applications

- Image-guided radiation therapy

Background and More Information

For better treatment guidance, cone beam CT (CBCT) scans are increasingly being used in image-guided radiation therapy (IGRT). However, the applications of CBCT are greatly hampered by its poor image quality due to scatter artifacts. In the current clinical practice, CBCT only provides patient geometry information for treatment setup. Quantitative x-ray CT images with high Hounsfield unit (HU) accuracy, which are particularly important for dose verification and adaptive radiation therapy, are still not achievable using CBCT. Many scatter correction algorithms have been proposed in literature, while a standard practical solution remains unclear.

Inventors

- Dr. Lei Zhu
Associate Professor - Georgia Tech School of Mechanical Engineering
- Tianye Niu
Research Scientist - Georgia Tech School of Mechanical Engineering

IP Status

:

Publications

, -

Images

Visit the Technology here:

[Scatter Correction for Cone-beam CT in Radiation Therapy](#)

<https://s3.sandbox.research.gatech.edu//index.php/print/pdf/node/3567>